


 Elcometer Limited  
 Edge Lane, Manchester  
 M43 6BU, England  
 Tel: +44 (0)161 371 6000  
 Fax: +44 (0)161 371 6010  
 sales@elcometer.com  
 www.elcometer.com  
 REGISTERED NO. 1729726 ENGLAND  
 Accredited to ISO 9001  
 ISO 14001

## Declaration of Conformity

### The Directives covered by this declaration:

2014/30/EU EMC Directive  
 2011/65/EU Restriction of the Use of certain Hazardous Substances in Electrical and Electronic Equipment (RoHS) Directive

Declares that the product(s): **Dakota CX2 & CX4 Ultrasonic Material Thickness Gauge**

Part Number(s): **CX2, CX4**

Product Option(s): **TXC10M0BP-1, TXC10M0CP-4, TXC1M00EP-2, TXC2M25CP-2, TXC2M25EP-2, TXC3M50EP-1, TXC5M00BP-4, TXC5M00CP-4, TXC5M00CP-6, TXC5M00CP-8, TXC5M00EP-3, TXC5M00EP-4, TXC7M50BP-3, TXC7M50CP-4, TXC7M50CP-5, T92024911, T99921325**

This declaration of conformity is issued under the sole responsibility of Elcometer Limited. The products identified above comply with the requirements of the above EU Directive by meeting the following standards:

<b>EN 61326-1:2013</b> <b>IEC 61326-1:2012</b> <b>Class B<sup>1</sup>, Group 1<sup>2</sup> ISM</b>	Electrical equipment for measurement, control and laboratory use – EMC requirements Part 1 General requirements.
<b>EN 61326-2-1:2013<sup>1</sup></b> <b>IEC 61326-2-1:2012</b>	Electrical equipment for measurement, control and laboratory use — EMC requirements Part 2-1: Particular requirements — Test configurations, operational conditions and performance criteria for sensitive test and measurement equipment for EMC unprotected applications
<b>EN 63000:2018</b>	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

<sup>1</sup> Class B product: Suitable for use in domestic establishments and in establishments directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

<sup>2</sup> Group 1 ISM product: Product in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.



Signed:



M. C. Sellars